



NQSUR

Urine – strip and sedimentation

Description:

Sent sample: 1 vial of liquid urine

Instruction of use:

Before testing bring the sample at room temperature (18-25 °C). For complete homogenization, invert the sample several times.

Analyze the control material as a patient sample.

Storage:

Store the sample at **2-8 °C**.

After opening the sample is stable **30 days** kept at **2-8 °C**

Precautions:

The urine has been tested for Hepatitis B Surface Antigen (HbsAg), and Hepatitis C Virus (HCV) antibody and found to be NON-REACTIVE. However this material should be handled carefully.

Parameters:

ascorbic acid, bilirubin, specific gravity, leukocytes, glucose, HCG, ketone, microalbumin, nitrite, protein, pH, urobilinogen, blood, urine sedimentation