

D-CHECK D
CONTROL
LOT: 1Z0601
ASSAY VALUES AND EXPECTED RANGES

Vizsgálati értékek és várható tartományok


05-12-2026

Instruments:		HemoCue [®]							
Parameter		CONTROL "Low"		CONTROL "Normal"		CONTROL "High"			
		LOT: 1Z0601 L		LOT: 1Z0601 N		LOT: 1Z0601 H			
		Mean	± Limit	Range Variation	Mean	± Limit	Range Variation	Mean	± Limit Range Variation
Hgb	g/dL	8,1	± 0,8	7,3 - 8,9	15,7	± 1,0	14,7 - 16,7	19,4	± 2,5 16,9 36,3
	g/L	81	± 8	73 - 89	157	± 10	147 - 167	194	± 25 169 363

